



# CERTIFICATE

Survey of 22 March 2017

You have fulfilled the requirements of the External Quality Assessment with the following analysis

## Hormones - Gravity - 1st Trimester Screening (305):

Validity 6 months:

free beta-chain hCG (R: B1a)

Validity 12 months:

PAPP-A

(R) analysis is subject to the RiliBÄK

**Participant:**

43733

CG Cytogenomics Medical Laboratory

CG Cytogenomics

Byzantioy 149

14233 Nea Ionia Athens

Düsseldorf, 21 April 2017

Prof. Dr. med. Michael Spannagl  
(Head of Reference Institution)



Prof. Dr. med. Friedrich Maly  
(Adviser)

Survey of 22 March 2017


Adviser:

Prof. Dr. med. Friedrich Maly  
Am Bärenkamp 22  
40589 Düsseldorf

Tel.: +49 211 602 1711  
Fax: +49 211 602 1713  
Mail: maly@instand-ev.de

**305**

## Hormones - Gravidity - 1st Trimester Screening

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Meets criteria	
free beta-chain hCG	21	U/l	40.9	40.3	SW	28.2	52.4	1.5%	+	
	22		9.55	8.99	SW	6.29	11.7	6.2%	+	
PAPP-A	21	U/l	3.23	3.21	SW	1.99	4.43	0.6%	+	
	22		80.6	80.2	SW	49.7	111	0.5%	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
free beta-chain hCG	21	ng/ml	40.9	1.00	236	HN	HN01
	22		9.55	1.00			
PAPP-A	21	mU/l	3226	0.001	236	HN	HN01
	22		80580	0.001			