

INSTAND e.V.  
Gesellschaft zur Förderung der Qualitätssicherung in medizinischen Laboratorien e. V.  
40223 Düsseldorf Ueber - Str. 20 , 40093 Postfach 250211, Germany

# Certificate

## Virus genome detection June 2016

Deadline of EQAS : 24.6.2016


**CG Cytogenomics Medical Laboratory (participant 43733 E)**  
**CG Cytogenomics**  
**Byzantioy 149**  
**14233 Nea Ionia Athens**  
**Greece**


This is to certify that the above mentioned participant has s u c c e s s f u l l y taken part in EQAS June 2016 with the following analyses:

PCR-/ NAT - HPV (373) - Differentiation of "High Risk" and "Low Risk" types of HPV (20) R-B3 \*\*  
PCR-/ NAT - HPV (373) - Typing of "High Risk" - HPV and "Low Risk" - HPV (30) R-B3 \*\*

\* valid for 6 months \*\* valid for 12 months \*\*\* valid for 2 years  
R-B2: RiliBÄK-parameter subject to B2 R-B3: RiliBÄK-parameter subject to B3



  
Prof./Dr. M. Spannagel  
(Chairman of the reference institution)

  
Prof. Dr. H. Zeichhardt  
(Expert)

Düsseldorf, 3.8.2016

Rv=1622 CvirZ / Lzert 3.8.2016 13:18h

**INSTAND EQAS**  
**Virus genome detection**
**Individual results**

 participant 43733  
 3. 8.2016  
 page 2 -E

June 2016

## Your results:

sample	test - method / reagent	evaluation of your results	your qual. result
<b>Differentiation of "High Risk" and "Low Risk" types of HPV (20)</b>			
373061	in house	correct	High Risk
373062	in house	correct	Low Risk
373063	in house	correct	High Risk
373064	in house	correct	negative
373065	in house	+	correct High Risk
<b>Typing of "High Risk" - HPV and "Low Risk" - HPV (30)</b>			
373061	in house	correct	type: 18
373062	in house	correct	type: 6
373063	in house	correct	type: 16
373064	in house	correct	negative
373065	in house	+	correct type: 18